



*Sample Buffer*  
*Sample Buffer without Tween*

(article no. 105)  
(article no. 106)

**Sample and antibody dilution buffer for immunoassays**

|                                      |   |
|--------------------------------------|---|
| Storage:                             | 2 – 8 °C or -15 to -30 °C<br>(tolerates repeated freezing and thawing cycles) |
| pH-value at 19.0 – 21.0 °C:          | 7.2 ± 0.2   |
| Preservative:                        | contains < 0.0014 % [w/w] reaction mass of CMIT/MIT (3:1)                     |
| Expiry date<br>when stored unopened: | see label on the bottle   |

**For general laboratory use**

**Instructions for use**

*Sample Buffer* is ready-to-use. Please shake the buffer thoroughly before use.

Standards and samples should be diluted with *Sample Buffer* at 1:2 or higher. A useful dilution for most applications is 1:10 (1 part sample in 9 parts *Sample Buffer*). Standards and samples should be treated identically.

The suitability of *Sample Buffer* for the respective assay must be tested by the user.

*Sample Buffer* is designed for trouble free assays without interference. If you have background or false-positives - for example derived from cross reactivities or matrix effects - we recommend using *LowCross-Buffer*<sup>®</sup> or *Assay Defender*<sup>®</sup>.

For further information please visit [www.candor-bioscience.com](http://www.candor-bioscience.com).

LowCross-Buffer and Assay Defender are registered trade marks of CANDOR Bioscience GmbH.